Thermometer Calibration Log (Calibration is the process of resetting a thermometer to read accurately at a known temperature)



Calibration Instructions:

- Fill a large glass with ice and water then stir.
- Immerse thermometer stem 2 inches in water until needle stops moving. Hold the stem away from the bottom and sides to avoid error.
- Rotate the nut *(under head)* until the pointer reads **32°F** while in ice water by using a mini wrench or thermometer cover adjusting hole.
- Calibrate weekly, when thermometer is dropped or when changing from one extreme temperature to another.
- Sanitize thermometers before using by wiping with an alcohol swab or by washing in soap, rinsing and sanitizing.

Date	Temp. Before Calibration	Temp. After Calibration	Calibrated by	Comments	Verified By

WASHOE COUNTY HEALTH DISTRICT
ENHANCING QUALITY OF LIFE

ENVIRONMENTAL HEALTH SERVICES
1001 East Ninth Street I P.O. Box 11130 I Reno, Nevada 89520
775-328-2434 I Fax: 775-328-6176 I washoecounty.us/health

Public Health
Prevent. Promote. Protect.